Tel: 400-968-6088 Fax: 010-56371281

Http://www.solarbio.com

巴氏染色试剂盒(EA65)

货号: G1614

规格: 4×100mL/4×500mL

保存: 室温, 避光保存, 有效期1年。

产品组成:

名称	4×100mL	4×500mL	保存
试剂(A):苏木素染色液	100mL	500mL	室温, 避光
试剂(B):蓝化液	100mL	500mL	室温
试剂(C):橘黄G6染色液	100mL	500mL	室温, 避光
试剂(D): EA65染色液	100mL	500mL	室温, 避光

产品介绍:

细胞学常规染色普遍使用巴氏(Papanicolaou)法。Papanicolaou Stain 最初仅用于检测阴道上皮雌激素水平以及生殖道念珠菌、滴虫等病原体。橘黄 G6 与 EA36 或 EA50 联用使用,可将胞浆染成颜色鲜明的绿色、蓝色和粉色。目前大多数实验室采用成品染液,所以每种染液应注意其改良后的最佳条件。最终胞浆染色应透明可见,核染色质应很容易辨别出来。目前改良的巴氏染色液含有多种离子,具有多色性染色效能。染色后胞质鲜艳、透明性好以及核膜、核仁、染色质结构清晰。细胞核染色液主要为 Harris 苏木素染液,细胞质染色液主要为 EA36 染液、EA50 染液。巴氏染色液用于细胞脱落标本,细胞核呈蓝色或黑色,角化鳞状细胞胞浆呈粉红或橘红色。

巴氏染色试剂盒(EA65)细胞质染液采用 EA65,细胞核染色采用自主研发的无毒改良型苏木素染色液, EA65 更适用于非妇科细胞学涂片染色。

自备材料:

固定液(如AF固定液)、系列乙醇、显微镜、盐酸乙醇分化液

操作步骤: (仅供参考)

- 1. 细胞涂片用95%乙醇固定10-15min。
- 2. 95%、85%、75%乙醇分别浸泡1min。
- 3. 蒸馏水或自来水浸泡或冲洗1min。
- 4. 苏木素染液染色5-10min。 蒸馏水冲洗2min。
- 5. 1%的盐酸乙醇分化液分化约4-5s或0.5%盐酸水溶液分化10s。
- 6. 自来水冲洗2min,蓝化液中蓝化2min,自来水冲洗2min。
- 7. 75%、85%、95%乙醇分别浸泡2min。
- 8. 橘黄G6染液染色2min。 95%乙醇冲洗1min。
- 9. EA65染色液染色3-5min。95%乙醇冲洗1min。
- 10. 无水乙醇(I)、(II)脱水各1min。
- 11. 二甲苯透明,中性树脂封片。

染色结果:

细胞核	蓝紫色或黑色
非角化细胞的胞质	淡蓝色或淡绿色
角化细胞的胞质	粉红或橘红色

注意事项:

- 1. 所有染液均需过滤,需经常更换染液。
- 2. 为了您的安全和健康,请穿实验服并戴一次性手套操作。

第1页,共2页

Papanicolaou EA65 Stain Kit

Cat: G1614

Size: 4×100mL /4×500mL

Storage: RT, avoid light, valid for 1 year.

Kit Components

Reagent	4×100mL	4×500mL	Storage
Reagent(A):Hematoxylin Solution	100mL	500mL	RT, avoid light
Reagent(B): Bluing Solution	100mL	500mL	RT
Reagent(C): Orange G Solution	100mL	500mL	RT, avoid light
Reagent(D): EA65 Solution	100mL	500mL	RT, avoid light

Introduction

Papanicolaou method is widely used in routine cytological staining. Papanicolaou stain is initially used to detect estrogen levels in vaginal epithelium and pathogens such as Candida and trichomonad in the genital tract. The combination of orange G6 and EA36 or EA50 can dye the cytoplasm into bright green, blue and pink. At present, most laboratories use finished dye solution, so we should pay attention to the best conditions of each dye solution after improvement. The final cytoplasmic staining should be transparent, and the nuclear chromatin should be easily distinguished.

At present, the improved pasteurization solution contains many kinds of ions, which has polychromatic dyeing efficiency. After staining, the cytoplasm is bright and transparent, and the structure of nuclear membrane, nucleolus and chromatin is clear. The nuclear staining solution is mainly Harris hematoxylin staining solution, and the cytoplasm staining solution is mainly EA36 staining solution and EA50 staining solution. Papanicolaou Stain is used for exfoliated cells. The nucleus is blue or black, and the cytoplasm of keratinized squamous cells is pink or orange red.

Papanicolaou EA65 Stain Kit use EA65 staining solution for cytoplasmic staining and non-toxic modified hematoxylin staining solution for nuclear staining. It is more suitable for non gynecological cytology smear staining.

Self Provided Materials

Fixative(like AF Fixative), Series of alcohol, Microscope, Acid alcohol differentiation solution

Protocol(for reference only)

- 1. For cell smear, fix with 95% alcohol for 10-15mins.
- 2. Rinse in 95%, 85%, 75% alcohol separately for 1min. Rinse with tap or distilled water for 1min.
- 3. Stain with Hematoxylin Solution for 5-10mins. Rinse with tap water for 2mins.
- 4. Differentiate with 1% acid alcohol differentiation solution for about 4-5s or 0.5% acid alcohol differentiation solution for about 10s.Rinse with tap water for 2mins.
- 5. Blue with Bluing Solution for 2mins. Rinse with tap water for 2mins.
- 6. Dehydrate in 75%, 85%, 95% alcohol separately for 2mins.
- 7. Stain with Orange G Solution for 2mins. Dehydrate in 95% alcohol for 2mins twice.
- 8. Stain with EA65 Solution for 2mins. Dehydrate in 95% alcohol for 2mins twice.
- 9. Dehydrate in absolute alcohol for 1min twice.
- 10. Transparent with xylene and seal with resinene.

Result

Nucleus	Blue Purple or Black	
Cytoplasm of non keratinized cells	Light Blue or Light Green	
Cytoplasm of keratinocytes	Pink or Orange	

Note

- 1. All dye solutions should be filtered before use and changed frequently.
- 2. For your safety and health, please wear experimental clothes and disposable gloves.